

Product Datasheet

MF-Human MMP-8(Matrix Metalloproteinase 8) ELISA Kit (orb1806692)



www.biorbyt.com

Description MF-Human MMP-8 (Matrix

Reactivity Human

Range 0.16-10ng/mL

Target MMP-8

Note For research use only

Application notes This ELISA kit uses the Sandwich-ELISA

principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human MMP-8.

Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody and biotinylated detection antibody specific for Human MMP-8. Then Avidin-Horseradish

Peroxidase (HRP) conjugate are added

successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human MMP-8, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzymesubstrate reaction is terminated by the addition of stop solution and the color

wavelength of 450 nm. The OD value is proportional to the concentration of Human MMP-8. You can calculate the concentration of Human MMP-8 in the samples by comparing the OD of the

turns yellow. The optical density (OD) is measured spectrophotometrically at a

samples to the standard curve.

Sample Types saliva Sample volume 25μL

Sensitivity 0.1 ng/mL

Expiration Date Please enquire.

