



## **Product Datasheet**

## Human AGC(Aggrecan) ELISA Kit (orb1807711)

**Description** Human AGC(Aggrecan) ELISA Kit

Reactivity Human

**Range** 0.16-10 ng/mL

Target AGC

**Note** For research use only

**Application notes** This ELISA kit uses the Sandwich-ELISA

principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human AGC. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human AGC and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each

micro plate well and incubated. Free

components are washed away. The substrate solution is added to each well. Only those wells that contain Human AGC, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD)

is measured spectrophotometrically at a wavelength of 450 nm  $\pm$  2 nm. The OD value is proportional to the concentration of Human AGC. You can calculate the concentration of Human AGC in the samples by comparing the OD of the samples to the standard curve.

**Sample Types** serum, plasma, Tissue homogenate and Other

biological samples

**Sensitivity** 0.10 ng/mL

**Expiration Date** Please enquire.

