

## Human HDV ELISA Kit

### Cat#: orb339948 (CoA)

<b>Code</b>	orb339948									
<b>Target Name</b>	hepatitis D virus(HDV) antibody(IgG)									
<b>Alternative Names</b>	N/A									
<b>Abbreviation</b>	HDV Ab (IgG)									
<b>Species</b>	Homo sapiens (Human)									
<b>Sample Types</b>	serum, plasma									
<b>Detection Range</b>	Request Information									
<b>Sensitivity</b>	Request Information									
<b>Assay Time</b>	1-5h									
<b>Sample Volume</b>	50-100ul									
<b>Detection Wavelength</b>	450 nm									
<b>Precision</b>	<p>Intra-assay Precision (Precision within an assay): CV%&lt;15%</p> <hr/> <p>Three samples of known concentration were tested twenty times on one plate to assess.</p> <hr/> <p>Inter-assay Precision (Precision between assays): CV%&lt;15%</p> <hr/> <p>Three samples of known concentration were tested in twenty assays to assess.</p>									
<b>Typical Data</b>	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Test parameter</th> <th style="text-align: left;">specification</th> <th style="text-align: left;">test result</th> </tr> </thead> <tbody> <tr> <td>Positive control</td> <td>&gt;0.6</td> <td>0.842</td> </tr> <tr> <td>Negative control</td> <td>&lt;0.1</td> <td>0.051</td> </tr> </tbody> </table>	Test parameter	specification	test result	Positive control	>0.6	0.842	Negative control	<0.1	0.051
Test parameter	specification	test result								
Positive control	>0.6	0.842								
Negative control	<0.1	0.051								

	<hr/> <b>Positive rate</b> 10, Positive    100% <hr/> <b>Negative rate</b> 10, Negative    100%
<b>Storage</b>	Store at 2-8°C. Please refer to protocol.

**Biorbyt Ltd.**

5 Orwell Furlong, Cowley Road, Cambridge, Cambridgeshire  
CB4 0WY, United Kingdom  
Email: info@biorbyt.com | Phone: +44 (0)1223 859 353 | Fax: +44(0)1223 280 240

**Biorbyt LLC**

1100 Corporate Square Drive, Helix Center, Suite 221, St Louis  
MO 63132, United States  
Email: info@biorbyt.com | Phone: +1 (415)-906-5211 | Fax: +1 (415) 651 8558