

## Human Anti-chlamydia trachomatis (CT-Ab-IgM) ELISA Kit

Cat #: orb1668287 (manual)

*For research use only. Not for use in diagnostic or therapeutic procedures.*

### INTENDED USE

For the qualitative detection of Human Anti-chlamydia trachomatis IgM antibody (CT-Ab-IgM) concentration in serum, plasma and other biological fluids.

### REAGENTS AND MATERIALS PROVIDED

Reagents	Quantity	Reagents	Quantity
Assay plate (96 Wells)	1	Instruction manual	1
Positive Control	1 x 1 mL	Negative Control	1 x 1 mL
HRP-Conjugate	1 x 11 mL	Sample Diluent	1 x 20 mL
Wash Buffer (30 x concentrate)	1 x 20 mL	Stop Solution	1 x 6 mL
Substrate A	1 x 6 mL	Substrate B	1 x 6 mL
Adhesive Films	4		

### MATERIALS REQUIRED BUT NOT SUPPLIED

Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 630 nm.

Precision single or multi-channel pipettes and disposable tips.

Deionized or distilled water.

Eppendorf Tubes for serial dilution samples.

Container for Wash Solution.

Absorbent paper for blotting the microtiter plate.

## STORAGE

Unopened kit	Store at 2 - 8°C. Do not use past kit expiration date.	
Opened/ Reconstituted Reagents	Coated assay plate	May be stored for up to 1 month at 2 - 8°C. Return unused wells to the foil pouch containing the desiccant pack, reseal along entire edge of zip-seal, and avoid the damp.
	Positive Control	May be stored for up to 1 month at 2 - 8°C.
	Negative Control	
	HRP-Conjugate	
	Sample Diluent	
	Wash Buffer	
	Substrate A	
	Substrate B	
	Stop Solution	

## INTRODUCTION

Chlamydia is one of the most common sexually transmitted infections, although it is often underrecognized. It can cause symptoms such as discharge, burning sensations, abnormal bleeding, pain, and nausea; however, these symptoms are easily confused with those of gonorrhea. In addition, chlamydia is asymptomatic in nearly three-quarters of women and about half of men. If left untreated, the infection can spread silently and eventually damage the fallopian tubes and ovaries, leading to permanent injury to the female reproductive system and, in some cases, infertility. In men, it can also cause epididymitis, which may result in sterility. Fortunately, chlamydia can be detected through affordable and reliable testing and is highly responsive to antibiotic treatment. However, diagnosis and treatment are often delayed until complications have already developed. Beyond its role as a sexually transmitted infection, Chlamydia trachomatis is also the causative agent of trachoma, the leading cause of preventable blindness worldwide. Repeated infections can cause the eyelids to turn inward, and without treatment, the resulting scarring can ultimately lead to vision loss.

## PRINCIPLE OF THE ASSAY

The microtiter plate provided in this kit has been pre-coated with CT-Ag. Samples are then added to the appropriate microtiter plate wells with a Horseradish Peroxidase (HRP)-conjugated CT-Ab-IgM and incubated. Then substrate solutions are added to each well. Only those wells that contain CT-Ab-IgM and enzyme-conjugated CT-Ab-IgM will exhibit a change in color. The enzyme- substrate reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm ± 2 nm.

## LIMITATIONS OF THE PROCEDURE

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

The kit should not be used beyond the expiration date on the kit label.

Do not mix or substitute reagents with those from other lots or sources.

It is important that the Calibrator Diluent selected for the standard curve be consistent with the samples being assayed.

If samples generate values higher than the highest standard, dilute the samples with the appropriate Calibrator Diluent and repeat the assay.

Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors have been tested in the ELISA Kit, the possibility of interference cannot be excluded.

## STABILITY

The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage conditions.

The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37 °C for 4 and 7 days and compare O.D. values of the kit kept at 37 °C with that of at recommended temperature.

### Note:

To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.

## SAMPLE COLLECTION AND STORAGE

- **Serum** Use a serum separator tube (SST) and allow samples to clot for two hours at room temperature or overnight at 2 - 8 °C before centrifugation for 15 minutes at 1000 × g. Remove serum and assay immediately or aliquot and store samples at ≤ - 20 °C. Avoid repeated freeze- thaw cycles.
- **Plasma** Collect plasma using EDTA, or heparin as an anticoagulant. Centrifuge for 15 minutes at 1000 × g at 2-8°C within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ - 20 °C. Avoid repeated freeze- thaw cycles.
- **Other biological fluids** Centrifuge samples for 20 minutes at 1000 × g. Remove particulates and assay immediately or store samples in aliquot at - 20 °C or -80 °C. Avoid repeated freeze/ thaw cycles.

### Note:

1. Samples to be used within 5 days may be stored at 2-8 °C, otherwise samples must be stored at - 20°C (≤ 1 month) or - 80 °C (≤ 2 months) to avoid loss of bioactivity and contamination.
2. Sample hemolysis will influence the result, so hemolytic specimen cannot be detected.
3. When performing the assay, bring samples to room temperature.

## SAMPLE PREPARATION

Human serum or plasma samples require at least a 101-fold dilution before test. A suggested 101-fold dilution can be achieved by adding 2 µL sample + 200 µL Sample Diluent, mix well, incubate for 20

minutes at 37°C. The recommended dilution factor is for reference only. The optimal dilution factor should be determined by users according to their particular experiments.

## REAGENT PREPARATION

**Bring all reagents to room temperature before use.**

Wash Buffer (1 x) - If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (30 x) into deionized or distilled water to prepare 600 mL of Wash Buffer (1 x).

## ASSAY PROCEDURE

**Bring all reagents and samples to room temperature before use. It is recommended that all samples, controls, and standards be assayed in duplicate.**

1. Prepare all reagents, working standards, and samples as directed in the previous sections.
2. Refer to the Assay Layout Sheet to determine the number of wells to be used and put any remaining wells and the desiccant back into the pouch and seal the ziploc, store unused wells at 2 - 8°C.
3. Set a Blank well without any solution. Add 100 µL of Negative Control, Positive Control and Sample per well. Incubate for 30 minutes at 37°C.
4. Aspirate each well and wash, repeating the process for a total of three washes. Wash by filling each well with Wash Buffer (250 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µL of HRP-Conjugate to each well (**Note: Do not add to Blank!**). Cover the microtiter plate with adhesive films. Incubate for 30 minutes at 37°C.
6. Aspirate each well and wash, repeating the process for a total of three washes. Wash by filling each well with Wash Buffer (250 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
7. Add 50 µL of Substrate A and 50 µL of Substrate B to each well, mix well. Incubate for 10 minutes at 37°C. Keeping the plate away from drafts and other temperature fluctuations in the dark.
8. Add 50 µL of Stop Solution to each well when the first four wells containing the highest concentration of standards develop obvious blue color. If color change does not appear uniform, gently tap the plate to ensure thorough mixing.
9. Determine the optical density of each well within 5 minutes, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. Subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

**\* Samples may require dilution. See Sample Preparation section.**

## CALCULATION OF RESULTS

For calculation of the human CT-Ab-IgM, compare the sample well with control.

1) Negative Control OD values must no more than 0.21.

If the average value of  $OD_{\text{negative}} < 0.10$ , calculate it as 0.10.

If the average value of  $0.10 < OD_{\text{negative}} < 0.21$ , In terms of actual OD value.

If one of the Negative Control OD values is higher than 0.21, discard it.

If more than two Negative Control OD values are higher than 0.21, repeat the test.

2) Positive Control OD Values must no less than 0.40.

If one of the Positive Control OD values less than 0.40, discard it.

If the two Positive Control OD value less than 0.40, repeat the test.

3) Cutoff Value= the average of Negative Control OD values  $\times 2.1$

While  $OD_{\text{sample}} / OD_{\text{negative}} \geq 2.1$ : Positive

While  $1.5 \leq OD_{\text{sample}} / OD_{\text{negative}} < 2.1$ : Suspicious

While  $OD_{\text{sample}} / OD_{\text{negative}} < 1.5$ : Negative

## IMPORTANT NOTE

1. There may be some foggy substance in the wells when the plate is opened at the first time. It will not have any effect on the final assay results. Do not remove microtiter plate from the storage bag until needed.
2. Do not mix or substitute reagents from one kit lot to another. Use only the reagents supplied by manufacturer.
3. Samples or reagents addition: Please use the freshly prepared Standard. Please carefully add samples to wells and mix gently to avoid foaming. Do not touch the well wall as far as possible. For each step in the procedure, total dispensing time for addition of reagents or samples to the assay plate should not exceed 10 minutes. This will ensure equal elapsed time for each pipetting step, without interruption. Duplication of all standards and specimens, although not required, is recommended. To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
4. Incubation: To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary. Do not allow wells to sit uncovered for extended periods between incubation steps. Once reagents have been added to the well strips, DO NOT let the strips DRY at any time during the assay. Incubation time and temperature must be observed.
5. Washing: The wash procedure is critical. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Solution by aspirating or decanting and remove any drop of water and fingerprint on the bottom of the plate. Insufficient washing will result in poor precision and falsely elevated absorbance reading. When using an automated plate washer, adding a 30 second soak period following the addition of wash buffer, and/ or rotating the plate 180 degrees between wash steps may improve assay precision.
6. Controlling of reaction time: Observe the change of color after adding Substrate Solution (e. g. observation once every 10 minutes), Substrate Solution should change from colorless or light blue to

Explore. Bioreagents.

gradations of blue. If the color is too deep, add Stop Solution in advance to avoid excessively strong reaction which will result in inaccurate absorbance reading.

7. Substrate Solution is easily contaminated. Substrate Solution should remain colorless or light blue until added to the plate. Please protect it from light.
8. Stop Solution should be added to the plate in the same order as the Substrate Solution. The color developed in the wells will turn from blue to yellow upon addition of the Stop Solution. Wells that are green in color indicate that the Stop Solution has not mixed thoroughly with the Substrate Solution.
9. Protect all reagents from strong light during storage and incubation. All the bottle caps of reagents should be covered tightly to prevent the evaporation and contamination of microorganism.
10. Wrong operations during the reagents preparation and loading, as well as incorrect parameter setting for the plate reader may lead to incorrect results. A microplate plate reader with a bandwidth of 10 nm or less and an optical density range of 0- 3 O. D. or greater at  $450 \pm 10$  nm wavelength is acceptable for **use** in absorbance measurement.
11. Even the same operator might get different results in two separate experiments. In order to get better reproducible results, the operation of every step in the assay should be controlled. Furthermore, a preliminary experiment before assay for each batch is recommended.
12. Limited by the current condition and scientific technology, we can't completely conduct the comprehensive identification and analysis on the raw material provided by suppliers. So there might be some qualitative and technical risks to use the kit.
13. The final experimental results will be closely related to validity of the products, operation skills of the end users and the experimental environments. Please make sure that sufficient samples are available.
14. Kits from different batches may be a little different in detection range, sensitivity and color developing time.
15. Each kit has been strictly passed Q. C test. However, results from end users might be inconsistent with our in-house data due to some unexpected transportation conditions or different lab equipments. Intra-assay variance among kits from different batches might arise from above factors, too.
16. Kits from different manufacturers with the same item might produce different results, since we haven't compared our products with other manufacturers.
17. The Stop Solution provided with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
18. Valid period: six months.