

# Taq Pro Multiple Probe qPCR Mix

QN213-EN



**Nanjing Vazyme Biotech Co.,Ltd.**

Tel: +86 25-83772625

Email: [info.biotech@vazyme.com](mailto:info.biotech@vazyme.com)

Web: [www.vazyme.com](http://www.vazyme.com)

Loc: Red Maple Hi-tech Industry Park, Nanjing, PRC

Follow Us



---

**Instruction for Use**

Version 22.1

01/Product Description ..... 02

02/Components..... 02

03/Storage ..... 02

04/Applications ..... 02

05/Notes ..... 03

06/Experiment Process (Take ABI 7500 as the test instrument) ..... 03

07/FAQ & Troubleshooting ..... 04

**01/Product Description**

Taq Pro Multiple Probe qPCR Mix is a master mix for probe qPCR to detect DNA templates (such as DNA viruses). The core component, Taq Pro HS DNA Polymerase, is a new generation hot start DNA polymerase that has been modified based on antibody technique and upgraded to improve template affinity. Equipped with the most suitable buffer optimized for the qPCR system, amplification performance with multiple targets, amplification specificity and sensitivity for detection of low-copy genes and amplification curve shape are significantly improved. It can provide excellent amplification curve within a wide quantitative range, and accurately quantify and detect target genes, with good repeatability and high reliability. It has wide compatibility in terms of template type, template GC content, and primer Tm values. And it also has good tolerance to impurities and is suitable for use in a variety of testing scenarios. The product is a 2 × master mix. Only primers, probes and templates need to be added additionally. It is convenient to use and compatible with fast program to reduce test time.

**02/Components**

Components	QN213-EN01 (100 rxns/20 µl reaction)	QN213-EN02 (500 rxns/20 µl reaction)	QN213-EN03 (2,500 rxns/20 µl reaction)
2 × Taq Pro Multiple Probe qPCR Mix <sup>a</sup>	1 ml	5 × 1 ml	
50 × ROX Reference Dye 1 <sup>b</sup>	100 µl	200 µl	5 × QN213-02
50 × ROX Reference Dye 2 <sup>b</sup>	100 µl	200 µl	

a. It includes dNTP Mix, Mg<sup>2+</sup>, Taq Pro HS DNA Polymerase, etc.

b. Used to rectify the error of fluorescence signals between different wells. Use 50 × ROX Reference Dye 1 for ABI 7900HT/7300 Real-Time PCR System and StepOnePlus; Use 50 × ROX Reference Dye 2 for ABI 7500, 7500 Fast Real-Time PCR System, and Stratagene Mx3000P. Don't use ROX for Roche and Bio-Rad Real-Time PCR instruments.

**03/Storage**

Store at -30 ~ -15°C and transport at ≤0°C.

**04/Applications**

This product is suitable for probe based qPCR detection of animal, plant and microbial DNA.

## 05/Notes

For research use only. Not for use in diagnostic procedures.

### Primer Design Guidance

1. The ideal primer length is 17 - 25 bp. Primers that are too short are likely to result in reduced amplification efficiency. Primers that are too long are more likely to generate higher structures. Both of them will interfere the accuracy of the quantitative results.
2. Control the GC content of the primers at 40% - 60%, and the optimum GC content is from 45% - 55%.
3. The  $T_m$  value of the primer should be greater than 60°C. Primer Premier 5 is recommended to calculate the  $T_m$  value.
4. The overall distribution of A, G, C, and T in the primer should be as uniform as possible. Avoid using regions with high GC or TA content, especially at the 3' end. Regions with uneven GC content must be avoided.
5. Try to avoid structures with consecutive T/C or A/G when designing primers.
6. The last five bases at the 3' end of the primer must not contain more than two G or C.
7. Ensure that the forward or the reverse primer is as close as possible to the probe sequence, but does not have overlap regions with the probe sequence.

### Guide for Designing TaqMan Probes

1. The probe sequence should be as close as possible to the forward or reverse primer, but overlap sequences must be avoided.
2. The probe length is typically 18 - 40 bp.
3. Avoid consecutive identical bases, especially four or more consecutive G.
4. Avoid using G at the 5' end of the probe.
5. The annealing temperature of the probe should be 65 ~ 67°C.
6. If the sequence contains polymorphic sites, the polymorphic sites should be located in the middle of the probe sequence.

## 06/Experiment Process (Take ABI 7500 as the test instrument)

### 1. Prepare the following mixture in a qPCR tube

2 × Taq Pro Multiple Probe qPCR Mix	10 $\mu$ l
Primer Mix	X $\mu$ l
TaqMan Probe Mix	Y $\mu$ l
50 × ROX Reference Dye 2	0.4 $\mu$ l
Template DNA	Z $\mu$ l
ddH <sub>2</sub> O	Up to 20 $\mu$ l

The amount of each component in the reaction system can be adjusted based on the following principles:

- ▲ The final concentration of primers in the reaction system is typically set at 0.2  $\mu$ M for better amplification results. If the reaction proceeds poorly, the primer concentration can be adjusted within a final concentration of 0.1 - 1.0  $\mu$ M.
- ▲ The final probe concentration can be adjusted to 50 - 250 nM.
- ▲ The sensitivity of qPCR is extremely high and the accuracy of the loading template amount can greatly affect the final quantitative results. It is recommended to dilute the template before adding, which can effectively improve experiment repeatability.
- ▲ Template can be adjusted to an appropriate volume. For example, if the template type is undiluted cDNA, then the volume should not exceed 1/10 of the total qPCR reaction volume.

### 2. Perform the qPCR reaction according to the following conditions

#### Standard Program

Stage 1	Contamination digestion	Rep: 1	37°C	2 min
Stage 2	Initial Denaturation	Rep: 1	95°C	30 sec
Stage 3	Cycles	Reps: 45	95°C	10 sec
			60°C	30 sec

#### Fast Program

Stage 1	Contamination digestion	Rep: 1	37°C	2 min
Stage 2	Initial Denaturation	Rep: 1	95°C	20 sec
Stage 3	Cycles	Reps: 45	95°C	1 sec
			60°C	20 sec*

\* Please confirm if the Real Time PCR instrument is in support of rapid amplification cycles.

## 07/FAQ & Troubleshooting

### ◇ Abnormal shape of amplification plot

- ① Rough amplification plot: The signal is too weak and generated after system correction. Increase template concentration and retry.
- ② Broken or downward amplification plot: The template concentration is too high and the baseline endpoint value is greater than  $C_T$  value. Reduce the baseline endpoint ( $C_T$  value - 4) and repeat data analysis.
- ③ Amplification plot goes downward suddenly: There are bubbles remaining in the reaction tube. Pay attention to centrifugation when processing samples and carefully check the reaction tube for any remaining bubbles before performing reaction.

### ◇ No amplification plot

- ① Insufficient number of reaction cycles: In general, the number of cycles is set to 40, but it should be noted that too many cycles will increase too many background signals and reduce the reliability of the data.
- ② Confirm whether the signal acquisition step is set up in the program: The two-step amplification program generally sets the signal acquisition at the annealing and extension stage, while the three-step amplification program should set the signal acquisition at the 72°C extension stage.

- ③ Confirm whether the primers are degraded: Primers that have not been used for a long time should be tested for integrity using PAGE electrophoresis before use in order to rule out the possibility of degradation.
- ④ Low template concentration: Reduce the dilution factor and repeat the test. In general, samples with unknown concentration should be started at the highest concentration.
- ⑤ Template degradation: Prepare new template and retry.

◇ **C<sub>T</sub> value appears too late**

- ① Low amplification efficiency: Optimize the PCR system, then try the three-step amplification program or redesign and synthesize primers.
- ② Low template concentration: Reduce the dilution factor and repeat the test. In general, samples with unknown concentration should be started at the highest concentration.
- ③ Template degradation: Prepare new template and retry.
- ④ Long PCR products: The recommended length of PCR product is 80 - 150 bp.
- ⑤ PCR inhibitors in the system: They are usually introduced along with the template. Increase the dilution factor or prepare new template and retry.

◇ **Amplification observed in negative control**

- ① Contaminated of reaction system: Replace with new mix, ddH<sub>2</sub>O and primers to repeat the experiment. The reaction system should be prepared in clean bench to reduce aerosol contamination.

◇ **The linear relationship of the standard curve is not satisfactory in absolute qualification**

- ① Deviations of pipetting volume: Increase the dilution factor of template and increase the pipetting volume accordingly.
- ② Degradation of standards: Prepare new standards and retry.
- ③ High template concentration: Increase the dilution factor.

◇ **Poor experiment repeatability**

- ① Inaccurate pipetting volume: Use higher performance pipette; increase the template dilution factor, and increase the sample loading volume.
- ② Differences in temperature control between wells in qPCR instrument: Calibrate the instrument regularly.
- ③ Low template concentration: The lower the template concentration, the worse the repeatability. Reduce the template dilution factor or increase the volume of sample addition.