

## **Product Datasheet**

## Human Soluble Interleukin 30 (sIL30) ELISA Kit (orb1088138)

Catalog Number orb1088138

Category Assays and Kits

**Description** The test principle applied in this kit is Sandwich enzyme immunoassay. The

microtiter plate provided in this kit has been pre-coated with an antibody specific to Soluble Interleukin 30(sIL30). Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific

to Soluble Interleukin 30(sIL30). Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Soluble Interleukin 30(sIL30), biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition

of sulphuric acid solution and the color change is measured

spectrophotometrically at a wavelength of  $450 \text{nm} \pm 10 \text{nm}$ . The concentration of Soluble Interleukin 30(sIL30) in the samples is then determined by comparing

the OD of the samples to the standard curve.

**Reactivity** Human

Concentration 500 pg/mL

**Application notes** standard: 500 pg/mL. Test principle: The test principle applied in this kit is

Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human sIL30. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human sIL30. Next, Avidin conjugated to Horseradish

Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human sIL30, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm  $\pm$  10nm. The concentration of Human sIL30 in the samples

is then determined by comparing the OD of the samples to the standard curve





**Assay Type** Sandwich

Assay Time 3.5h

**Range** 7.82-500 pg/mL


**Sensitivity** 3 pg/mL

**Sample Types** Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and

other biological fluids.

**Note** For research use only

**Expiration Date** Please enquire.



 $\begin{aligned} & \text{Email: } \underline{info@biorbyt.com}, \ \underline{support@biorbyt.com} \\ & \text{Phone: } \underline{+1 \ (415) \ 906-5211} \ \big| \ \text{Fax: } \underline{+1 \ (415) \ 651-8558} \end{aligned}$