

Product Datasheet SARS-CoV-2 (COVID-19) Spike S1 Antibody (orb1239995)

Catalog Number orb1239995

Category Antibodies

Description SARS-CoV-2 (COVID-19) Spike S1 Antibody

Target S

Clonality Polyclonal

Isotype IgG

Conjugation Unconjugated

Reactivity Virus

Form/Appearance Liquid

Concentration 1 mg/mL

Buffer/Preservatives SARS-CoV-2 (COVID-19) Spike S1 antibody is supplied in PBS containing 0.02%

sodium azide.

Purification SARS-CoV-2 (COVID-19) Spike S1 antibody is affinity chromatography purified via

peptide column.

Immunogen Anti-SARS-CoV-2 (COVID-19) Spike S1 antibody (orb1239995) was raised against

a peptide corresponding to 16 amino acids near the amino terminus of SARS-CoV-2 (COVID-19) Spike S1 glycoprotein. The immunogen is located within the

first 50 amino acids of SARS-CoV-2 (COVID-19) Spike S1 protein.

UniProt ID PODTC2

Tested applications ELISA, ICC, IF, IHC, WB





Antibody Type Primary Antibody


Storage Maintain refrigerated at 2-8°C for up to 2 weeks. For long term storage store at -

20°C in small aliquots to prevent freeze-thaw cycles.


Note For research use only

NCBI QHD43416

Expiration Date 12 months from date of receipt.




Immunohistochemistry Validation of Spike in the Nasal Swab Sample of Omicron Variant COVID-19 Patient.
Immunohistochemical analysis of 4% paraformaldehyde-fixed patient nasal swab sample using spike S1 antibody (orb1239995) at 0.5 µg/ml. Tissue was fixed with formaldehyde and blocked with 10% serum for 1 h at RT; antigen retrieval was by heat mediation with a citrate buffer (pH6). Samples were incubated with primary antibody overnight at 4°C. A goat antirabbit IgG H&L (HRP) at 1/250 was used as secondary. Counter stained with Hematoxylin. Strong spike protein signal was observed in the nasal swab sample of the Omicron variant COVID-19 patient.




Immunohistochemistry Validation of Spike in Delta Variant COVID-19 Patient Lung Tissue. Immunohistochemical analysis of paraffin-embedded patient lung tissue using anti-Spike S1 antibody (orb1239995) at 0.5 µg/ml. Tissue was fixed with formaldehyde and blocked with 10% serum for 1 h at RT; antigen retrieval was by heat mediation with a citrate buffer (pH6). Samples were incubated with primary antibody overnight at 4°C. A goat anti-rabbit IgG H&L (HRP) at 1/250 was used as secondary. Counter stained with Hematoxylin. Strong spike protein signal was observed in the Delta variant COVID-19 patient lung, but not in non-COVID-19 patient lung.








Immunohistochemistry Validation of SARS-CoV-2 (COVID-19) Spike S1 in Human Lung Tissue from the COVID-19 Patient (Sun et al., 2020). Detection of SARS-CoV-2 nucleocapsid protein by anti-SARS-COV-2 nucleocapsid antibodies (orb1239972, 0.02 µg/mL, A, B) or SARS-CoV-2 Spike S1 antibodies (orb1239995, 1 µg/mL D, E) in adjacent sections of autopsy lung tissue from COVID-19 deceased patient. Negative control staining on autopsy lung tissue from patient who died from non-COVID-19 pneumonia is shown for Nucleocapsid protein (C) or Spike protein (F). Negative control using normal rabbit immunoglobulin on COVID-19 autopsy tissue is presented (G). DAB chromogen and hematoxylin counterstain are used. Scale bars: 50µM in A, C, D, F, G; 20µM in B and E.



Immunohistochemistry Validation of SARS-CoV-2 (COVID-19) Spike S1 in COVID-19 Patient Lung. Immunohistochemical analysis of paraffin-embedded COVID-19 patient lung tissue using anti-SARS-CoV-2 (COVID-19) Spike S1 antibody (orb1239995, 0.5 μ g/mL). Tissue was fixed with formaldehyde and blocked with 10% serum for 1 h at RT; antigen retrieval was by heat mediation with a citrate buffer (pH6). Samples were incubated with primary antibody overnight at 4 °C. A goat antirabbit IgG H&L (HRP) at 1/250 was used as secondary. Counter stained with Hematoxylin. Strong signal of SARS-COV-2 spike protein was observed in macrophage of COVID-19 patient lung, but not in non-COVID-19 patient lung.



Immunohistochemistry Validation of SARS-CoV-2 (COVID-19) Spike S1 in Human Lung Tissue from the COVID-19 Patient. Immunohistochemical analysis of paraffin-embedded COVID-19 patient lung tissue using anti-SARS-CoV-2 (COVID-19) Spike S1 antibody (orb1239995, 1 μ g/mL). Tissue was fixed with formaldehyde and blocked with 10% serum for 1 h at RT; antigen retrieval was by heat mediation with a citrate buffer (pH6). Samples were incubated with primary antibody overnight at 4°C. A goat anti-rabbit IgG H&L (HRP) at 1/250 was used as secondary. Counter stained with Hematoxylin. (Courtesy of Hallgeir Rui, MCW) (Picture shown in 40X magnification).


7 Signet Court, Swann's Road, Cambridge, CB5 8LA, United Kingdom Email: info@biorbyt.com Phone: +44 (0) 1223 859-353 | Fax: +1 (415) 651-8558

Biorbyt LLC.


68 TW Alexander Drive,
Durham, NC, 27713, United States
Email: info@biorbyt.com
Phone: +1 (415) 906-5211 | Fax: +1 (415) 651-8558








Overexpression Validation in Spike Transfected 293 Cells. Loading: 10 μ g per lane of 293 cell lysate. Antibodies: SARS-CoV-2 (COVID-19) Spike S1, orb1239995 (1 μ g/mL), 1h incubation at RT in 5% NFDM/TBST. Secondary: Goat anti-rabbit lgG HRP conjugate at 1:10000 dilution. Lane 1: WT 293 cells and Lane 2: SARS-CoV-2 Spike overexpressed 293 cells.




Immunofluorescence Validation of SARS-CoV-2 (COVID-19) Spike S1 in 293 Transfected Cells. Immunofluorescent analysis of 4% paraformaldehyde-fixed Spike transfected 293 cells labeling SARS-CoV-2 (COVID-19) Spike S1 with orb1239995 at 5 ug/mL, followed by goat anti-rabbit IgG secondary antibody at 1/500 dilution (green) and DAPI staining (blue).



ELISA Validation with SARS-CoV-2 (COVID-19) Spike Recombinant Protein. Antibodies: SARS-CoV-2 (COVID-19) Spike S1 antibody, orb1239995. A direct ELISA was performed using SARS-CoV-2 (COVID-19) Spike S1 recombinant protein as coating antigen and the anti-SARS-CoV-2 (COVID-19) Spike S1 antibody as the capture antibody. Secondary: Goat anti-rabbit IgG HRP conjugate at 1:20000 dilution. Detection range is from 8 ng/mL to 1000 ng/mL.







Western Blot Validation with SARS-CoV-2 (COVID-19) Spike Recombinant Protein. Loading: 50 ng per lane of SARS-CoV-2 (COVID-19) Spike S1 recombinant protein. Antibodies: SARS-CoV-2 (COVID-19) Spike 1, orb1239995, 1h incubation at RT in 5% NFDM/TBST. Secondary: Goat anti-rabbit IgG HRP conjugate at 1:10000 dilution. Lane 1: 1 μ g/mL and Lane 2: 2 μ g/mL.

Phone: <u>+1 (415) 906-5211</u> | Fax: <u>+1 (415) 651-8558</u>