

Product Datasheet

Mouse Maresin 1 (MaR1) ELISA Kit (orb1817289)

Catalog Number orb1817289

Description This assay employs the competitive inhibition enzyme immunoassay technique.

The microtiter plate provided in this kit has been pre-coated with Mouse MaR1 protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse MaR1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm \pm 10nm.

The concentration of Mouse MaR1 in the samples is then determined by

comparing the OD of the samples to the standard curve.

Reactivity Mouse

Range 15.63-1000 pg/mL

Concentration 1000 pg/mL

Note For research use only

Application notes standard: 1000 pg/mL. Test principle: This assay employs the competitive

inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with Mouse MaR1 protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Mouse MaR1. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the

addition of sulphuric acid solution and the color change is measured

spectrophotometrically at a wavelength of $450 \text{nm} \pm 10 \text{nm}$. The concentration of Mouse MaR1 in the samples is then determined by comparing the OD of the

samples to the standard curve

Sample Types serum, plasma, tissue homogenates, cell lysates, cell culture supernates and

other biological fluids






Assay Time 2.5h

Sensitivity 6.06 pg/mL

Expiration Date Please enquire.



Email: <u>info@biorbyt.com</u>, <u>support@biorbyt.com</u>
Phone: <u>+1 (415) 906-5211</u> | Fax: <u>+1 (415) 651-8558</u>