

## Product Datasheet

### General Protein A (PROA) ELISA Kit (orb2307985)

<b>Catalog Number</b>	orb2307985
<b>Description</b>	<p>This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with PROA protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to PROA. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of <math>450\text{nm} \pm 10\text{nm}</math>. The concentration of PROA in the samples is then determined by comparing the OD of the samples to the standard curve.</p>
<b>Reactivity</b>	All
<b>Range</b>	46.88-3000 pg/mL
<b>Concentration</b>	3000 pg/mL
<b>Note</b>	For research use only
<b>Application notes</b>	<p>standard: 3000 pg/mL. Test principle: This assay employs the competitive inhibition enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with PROA protein. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to PROA. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of <math>450\text{nm} \pm 10\text{nm}</math>. The concentration of PROA in the samples is then determined by comparing the OD of the samples to the standard curve</p>
<b>Sample Types</b>	serum, plasma, other biological fluids
<b>Assay Time</b>	2.5h

#### Biorbyt Ltd.

7 Signet Court, Swann's Road,  
Cambridge, CB5 8LA, United Kingdom  
Email: [info@biorbyt.com](mailto:info@biorbyt.com), [support@biorbyt.com](mailto:support@biorbyt.com)  
Phone: [+44 \(0\) 1223 859-353](tel:+44(0)1223859353) | Fax: [+1 \(415\) 651-8558](tel:+1(415)6518558)

#### Biorbyt LLC.

68 TW Alexander Drive,  
Durham, NC, 27713, United States  
Email: [info@biorbyt.com](mailto:info@biorbyt.com), [support@biorbyt.com](mailto:support@biorbyt.com)  
Phone: [+1 \(415\) 906-5211](tel:+1(415)9065211) | Fax: [+1 \(415\) 651-8558](tel:+1(415)6518558)

<b>Sensitivity</b>	19.1 pg/mL
<b>Expiration Date</b>	Please enquire.

---

**Biorbyt Ltd.**

7 Signet Court, Swann's Road,  
Cambridge, CB5 8LA, United Kingdom  
Email: [info@biorbyt.com](mailto:info@biorbyt.com), [support@biorbyt.com](mailto:support@biorbyt.com)  
Phone: [+44 \(0\) 1223 859-353](tel:+44(0)1223859353) | Fax: [+1 \(415\) 651-8558](tel:+1(415)6518558)

**Biorbyt LLC.**

68 TW Alexander Drive,  
Durham, NC, 27713, United States  
Email: [info@biorbyt.com](mailto:info@biorbyt.com), [support@biorbyt.com](mailto:support@biorbyt.com)  
Phone: [+1 \(415\) 906-5211](tel:+1(415)9065211) | Fax: [+1 \(415\) 651-8558](tel:+1(415)6518558)