

Product Datasheet

Direct qPCR ProbesMaster (orb2646470)

Catalog Number orb2646470

Category Tools

Description Direct qPCR ProbesMaster is designed for quantitative real-time analysis of target DNA directly from blood, swabs and animal- or plant tissue. The mix allows robust amplification avoiding the requirement of any prior DNA purification procedures. The mix is recommended for use with dual-labeled fluorescent probes, e.g. TaqMan, Molecular Beacon or Scorpion probes. It provides a powerful tool for multiplex-quantification of sample DNA in a broad dynamic range with exceptional sensitivity and precision. The mix contains all reagents required for qPCR (except template, primer and labeled fluorescent probe) in a premixed 2 x concentrated ready-to-use solution. High robustness, reliability and sensitivity of the mix are based on an antibody-blocked hot start polymerase in combination with an optimized and well-balanced buffer system. The mix ensures fast and easy preparation with a minimum of pipetting steps and is specially recommended for:
Interference of remaining components from sample matrix
Due to the fast and easy but relatively rough sample preparation remaining components from the sample matrix may be co-transferred into the PCR assay. These remains are mostly blocked by a combination of specially optimized additives. If inhibition of the PCR reaction occurs with higher volumes of transferred sample volume, please reduce the sample volumes or use a dilution of the sample in 1x Direct Extraction Buffer or water.
Remaining components of the sample matrix may also show fluorescence signals specially in the yellow and red spectral range. If using this fluorescence range for multiplex PCR assays or if using ROX, please take special attention that there is no interference between fluorescence from remaining matrix material and the used fluorescence channels for amplicon detection.
ROX Reference Dye
The mix can also be used in combination with ROX reference dye (#PCR-351) in PCR instruments that are compatible with the evaluation of the ROX signal.

Form/Appearance liquid

Biorbyt Ltd.

7 Signet Court, Swann Road
Cambridge
CB5 8LA
United Kingdom

Email: info@biorbyt.com, support@biorbyt.com
Phone: [+44 \(0\)1223 859353](tel:+44(0)1223859353) | Fax: [+1 \(415\) 651-8558](tel:+1(415)651-8558)

Biorbyt LLC

68 TW Alexander Drive
Research Triangle Park
Durham
NC 27713
United States

Email: info@biorbyt.com, support@biorbyt.com
Phone: [+1 \(415\) 906-5211](tel:+1(415)906-5211) | Fax: [+1 \(415\) 651-8558](tel:+1(415)651-8558)

Concentration	2x
Storage	store at -20 °C. avoid freeze/thaw cycles. stable at 4 °C for up to 4 weeks
Note	For research use only
Expiration Date	12 months from date of receipt.

Biorbyt Ltd.

7 Signet Court, Swann Road
Cambridge
CB5 8LA
United Kingdom

Email: info@biorbyt.com, support@biorbyt.com

Phone: [+44 \(0\)1223 859353](tel:+44(0)1223859353) | Fax: [+1 \(415\) 651-8558](tel:+1(415)651-8558)

Biorbyt LLC

68 TW Alexander Drive
Research Triangle Park
Durham
NC 27713
United States

Email: info@biorbyt.com, support@biorbyt.com

Phone: [+1 \(415\) 906-5211](tel:+1(415)906-5211) | Fax: [+1 \(415\) 651-8558](tel:+1(415)651-8558)